

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)  
CB SCHEME

## CB TEST CERTIFICATE

Product

Switching power supply

Name and address of the applicant

N2POWER, DIVISION OF QUALSTAR  
130 W COCHRAN ST, SUITE C,  
SIMI VALLEY, CALIFORNIA 93065,  
USA

Name and address of the manufacturer

N2POWER, DIVISION OF QUALSTAR  
130 W COCHRAN ST, SUITE C,  
SIMI VALLEY, CALIFORNIA 93065,  
USA

Name and address of the factory

N2POWER, DIVISION OF QUALSTAR  
130 W COCHRAN ST, SUITE C,  
SIMI VALLEY, CALIFORNIA 93065,  
USA

Note: When more than one factory, please report on page 2

Ratings and principal characteristics

 Additional Information on page 2  
Input: 100-240 Vac, 50/60 Hz, 1.6 A Max.  
Output Rating: See test report for details.

Trademark (if any)



Type of Customer's Testing Facility (CTF) Stage used

Model / Type Ref.

ML100-X, MLE100-X, MLU100-X  
See Page 2

Additional information (if necessary may also be reported on page 2)

 Additional Information on page 2

A sample of the product was tested and found to be in conformity with

IEC 60601-1(ed.3), IEC 60601-1(ed.3);am1

As shown in the Test Report Ref. No. which forms part of this Certificate

170203701 issued on 2017-03-29

This CB Test Certificate is issued by the National Certification Body



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Denko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2017-03-30

Signature:

Jan-Erik Storgaard



Ref. Certif. No.

**DK-62409-UL**

**Model Details:**

ML100-X (X can be 12, 15, 18, 24, 28, 36 and 48);  
MLU100-X (X can be 12, 15, 18, 24, 28, 36 and 48);  
MLE100-X (X can be 12, 15, 18, 24, 28, 36 and 48)

**Additional Information:**

Additionally evaluated to EN 60601-1:2006 / A1:2013 / A12:2014; National Differences specified in the CB Test Report.  
Risk Management was not included in this investigation.

**Additional information (if necessary)**



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2017-03-30

Signature:

Jan-Erik Storgaard